

Our products are known for their high quality and durability.

This RC product complies with the provisions of the European Medical Devices Regulation 2017/745 „EU-MDR“.

Please read this User manual thoroughly and carefully before attempting to use this product and heed the given care instructions.

Familiarize yourself thoroughly with the operation of this unit before using it.

Retain this User's Manual for future use and pass it on to the next user of this device.

Use:

This diagnostic light is only to be used by authorized professionals.

Intended use:

The diagnostic penlights are used during procedures to illuminate areas of the body to be examined such as, for example, inside the mouth or the surface of the skin. Approved for eye examination (pupillary reflex).

(Photobiologically safety tested in accordance with standard DIN EN 62471:2009-03 for lamps and lamp systems).

Operation:

Remove the red transport isolation from the clip. To turn on, press the clip against the shaft of the pen until it is made with the metal ring on the shaft. To replace the batteries, unscrew the metal lamp holder. Insert new batteries with the positive pole facing the lamp.

Unsuitable use/contraindication of the products:

Any use other than that described here is not in accordance with the intended use. The manufacturer is not liable for any resulting damages.

**Warnings:**

Never allow the penlights to come into contact with fluids. No direct contact with the tongue or mucus membranes.

When examining the mouth cavity ensure that the light does not fall into the mouth due to the risk of suffocation!

If the lights show visible signs of damage, they are to be removed from clinical use.

Do not store near heating elements, not expose to direct sunlight. Higher temperatures can shorten the lifespan of the batteries.

Only original replacement lamps may be used, otherwise there is a risk of eye damage.

If the lights are visibly damaged, they must be removed from clinical use.

**Storage and transport:**

To avoid contamination and dust, please always keep the penlights in the packaging.

Storage preferably cool (below 30 °C) and dry, without large temperature fluctuations.

Caution: if the lamp falls, there is a risk of breakage (only for articles with a light bulb).

Cleaning:

The penlights are cleaned with a damp cloth on the surface.

Wipe disinfection:

The penlights can be disinfected with a cloth dampened with alcohol on the surface.

Use DGHM-tested alcohol-based cleaning wipes.

Reprocessing:

It is recommended that the instruments be cleaned as soon as possible after being used.

Sterilization:

Sterilization not possible.

Control and functional test:

Visually inspect the lamp before use. Test brightness, replace batteries if necessary.



Richard Cermak Miniaturlampen, Maximilianstr. 16a, 75172 Pforzheim



Separate collection of electrical and electronic items.
Used batteries are to be sent to a separate collection.